

TECHNICAL SPECIFICATIONS OF THE FDI France MEDICAL _TRIPOD_ Accessory for the elbow crutches or walking sticks



Definition: Accessory to allow elbow crutch to stand on its own

Internal reference nr.: TR/05/07

Structure: Made of polypropylene
With a central base (ferrule) to support the total weight of the user.
On the six extremities: equipped with small rubber bases

Base (Ferrule): Flexible design for maximum contact and adherence

- Equipped inside with an anti-perforation steel ring
- Replaceable

Weight: 110g

Height: 160mm

Central opening: 19mm

Width at the bottom: 150mm

Max. weight of user: Function of the maximum weight supported by the elbow crutch or the walking stick attached to the Tripod

Usage: Lower tube of the elbow crutch or walking stick to be inserted in central opening
TRIPOD allows elbow crutch or walking stick to stand on its own
TRIPOD remains attached to the elbow crutch or walking stick when it is used

Caution: During the walk, the TRIPOD must be in the up position. The support must be only on the base (ferrule) of the elbow crutch or the walking stick.

Label: CE mark
Lot number
Date of production
Name and address of manufacturer
Name and address of dealer, importer

Other Indications: CE mark engraved in TRIPOD

Manufacturing site: FDI FRANCE MEDICAL
77 Route de Buatière- Fitolieu
38490 Les Abrets en Dauphiné-France

Conformity: Medical Device Class I according EU Directive 93/42/EEC

